
Midmark M11 - v1.0.0
Total Cycles: 4480

Name: RITTER 1
Sterilizer ID: V2178198
04/23/2024 09:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.4 F 0.0
2:00 152.6 F 2.2
4:00 157.1 F 5.2
6:00 185.2 F 9.2
8:00 202.6 F 7.6
10:00 222.3 F 6.7
12:00 235.0 F 11.5
14:00 231.5 F 7.7
16:00 244.4 F 13.4
18:00 242.6 F 11.7
20:00 254.8 F 17.9
22:00 265.8 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.3 F 29.1
2:00 272.0 F 28.5
3:00 272.2 F 29.0
4:00 272.3 F 28.8

Min 271.4 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 224.7 F 0.0
10:00 233.7 F 0.0
15:00 229.5 F 0.0
20:00 231.9 F 0.0
25:00 225.5 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 23:38
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:01:54
CYCLE COMPLETE

04/23/2024 10:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____