

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
2.6uS / 1.6ppm  
CYCLE NUMBER 001017  
13:20 25/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:06  
PRESSURIZING 2:13  
STERILIZING 3:09  
136.0 C 228kPa 3:09  
136.0 C 229kPa 3:39  
136.3 C 232kPa 4:09  
136.3 C 232kPa 4:39  
136.2 C 230kPa 5:09  
136.4 C 232kPa 5:39  
136.3 C 231kPa 6:09  
136.4 C 232kPa 6:39  
136.4 C 232kPa 6:39  
VENTING 6:39  
AIR DRYING 7:27  
CYCLE COMPLETE 8:09

-----  
Digital Signature #  
B82F57661F16119D  
-----