
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006417
13:30 05/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:50
PRESSURIZING 1:56
STERILIZING 2:59
136.4 C 220kPa 2:59
136.0 C 222kPa 3:29
136.2 C 223kPa 3:59
136.0 C 221kPa 4:29
136.1 C 223kPa 4:59
136.4 C 220kPa 5:29
136.2 C 223kPa 5:59
136.2 C 222kPa 6:29
136.2 C 222kPa 6:29
VENTING 6:29
AIR DRYING 7:20
CYCLE COMPLETE 35:42

Digital Signature #
897FC3BBDBFE4361
