
Midmark M11 - v1.0.5
Total Cycles: 1384

Name: RITTER2
Sterilizer ID: V2313228
08/09/2022 11:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.1 F 0.0
2:00 167.8 F 0.1
4:00 184.3 F 2.9
6:00 207.2 F 7.7
8:00 226.4 F 13.0
10:00 231.2 F 9.2
12:00 237.5 F 9.7
14:00 244.1 F 13.1
16:00 243.3 F 12.3
18:00 256.2 F 19.1
20:00 267.5 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.6 F 29.8
2:00 272.7 F 29.4
3:00 271.5 F 28.9
4:00 271.4 F 29.1

Min 271.2 F 28.7
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.4 F 0.0
5:00 234.0 F 0.0
10:00 230.4 F 0.0
15:00 221.4 F 0.0
20:00 218.4 F 0.0
25:00 220.6 F 0.0
30:00 217.1 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:26
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:42
CYCLE COMPLETE

08/09/2022 12:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____