
Midmark M11 - v1.0.5
Total Cycles: 3470

Name: RITTER2
Sterilizer ID: V2313228
03/22/2024 10:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.8 F 0.0
2:00 117.9 F 0.0
4:00 130.8 F 0.6
6:00 157.1 F 3.2
8:00 175.5 F 6.8
10:00 194.1 F 10.8
12:00 214.1 F 5.1
14:00 222.5 F 8.4
16:00 231.3 F 12.9
18:00 235.4 F 9.8
20:00 238.2 F 9.5
22:00 247.1 F 14.1
24:00 258.4 F 20.2
26:00 268.6 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.3 F 29.7
2:00 271.6 F 29.1
3:00 272.6 F 29.6
4:00 271.7 F 29.5

Min 270.9 F 28.6
Max 273.2 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.1 F 0.0
5:00 196.5 F 0.0
10:00 195.2 F 0.0
15:00 195.5 F 0.0
20:00 193.0 F 0.0
25:00 190.4 F 0.0
30:00 188.0 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 27:04
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:05:29
CYCLE COMPLETE

03/22/2024 11:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____