
Midmark M11 - v1.0.5
Total Cycles: 655

Name: RITTER2
Sterilizer ID: V2313228
12/20/2021 02:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.8 F 0.0
2:00 176.4 F 2.8
4:00 186.6 F 7.6
6:00 213.7 F 12.9
8:00 223.4 F 7.9
10:00 237.8 F 13.3
12:00 239.5 F 11.4
14:00 239.1 F 10.5
16:00 253.3 F 17.6
18:00 265.4 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.5 F 29.5
2:00 271.3 F 29.0
3:00 272.4 F 29.6
4:00 271.5 F 28.9

Min 271.3 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.9 F 0.0
5:00 215.2 F 0.0
10:00 239.4 F 0.0
15:00 228.1 F 0.0
20:00 216.8 F 0.0
25:00 216.5 F 0.0
30:00 210.6 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 19:44
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:22
CYCLE COMPLETE

12/20/2021 03:53 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____