
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001001
14:11 16/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:33
PRESSURIZING 0:42
STERILIZING 1:35
136.0 C 221kPa 1:35
136.2 C 223kPa 2:05
136.4 C 223kPa 2:35
136.2 C 224kPa 3:05
136.3 C 222kPa 3:35
136.4 C 225kPa 4:05
136.4 C 223kPa 4:35
136.4 C 222kPa 5:05
136.4 C 222kPa 5:05
VENTING 5:05
AIR DRYING 5:59
CYCLE COMPLETE 6:02

Digital Signature #
7401D012921700A8
