

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3311

Name: RITTER2  
Sterilizer ID: V2313228  
01/19/2024 02:37 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.0 F 0.0  
2:00 161.9 F 2.2  
4:00 186.5 F 6.7  
6:00 206.3 F 12.3  
8:00 222.6 F 7.3  
10:00 236.0 F 13.1  
12:00 239.0 F 11.0  
14:00 238.5 F 10.0  
16:00 253.6 F 17.5  
18:00 266.3 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.9  
1:00 272.2 F 29.0  
2:00 272.8 F 29.5  
3:00 271.9 F 29.1  
4:00 272.5 F 29.2  
  
Min 271.7 F 28.7  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.3 F 0.0  
5:00 215.0 F 0.0  
10:00 229.8 F 0.0  
15:00 228.6 F 0.0  
20:00 225.0 F 0.0  
25:00 224.5 F 0.0  
30:00 221.9 F 0.0  
Drying Complete

FILLING: 1:42  
HEATING: 19:26  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:18  
CYCLE COMPLETE

01/19/2024 03:35 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_