
Midmark M11 - v1.0.0
Total Cycles: 1305

Name: RITTER 1
Sterilizer ID: V2178198
10/27/2021 08:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.5 F 0.0
2:00 168.1 F 1.3
4:00 172.7 F 3.9
6:00 191.1 F 7.6
8:00 210.6 F 12.2
10:00 220.3 F 5.9
12:00 232.8 F 10.7
14:00 233.3 F 7.5
16:00 244.7 F 13.3
18:00 243.5 F 11.9
20:00 256.0 F 18.5
22:00 267.1 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.7 F 29.0
2:00 271.5 F 28.2
3:00 272.9 F 29.1
4:00 271.9 F 28.4

Min 271.5 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.1 F 0.0
5:00 218.8 F 0.0
10:00 212.9 F 0.0
15:00 219.5 F 0.0
20:00 224.4 F 0.0
25:00 227.2 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 23:29
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:01:58
CYCLE COMPLETE

10/27/2021 09:11 AM
ACCEPTED: X
REJECTED:
Operator: U01