
Midmark M11 - v1.0.0
Total Cycles: 4944

Name: RITTER 1
Sterilizer ID: V2178198
09/30/2024 12:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.3 F 0.0
2:00 76.5 F 0.0
4:00 87.3 F 0.0
6:00 124.3 F 0.2
8:00 155.3 F 0.5
10:00 182.1 F 0.9
12:00 204.8 F 4.2
14:00 211.8 F 7.5
16:00 223.8 F 11.2
18:00 224.6 F 5.6
20:00 236.9 F 10.6
22:00 234.3 F 7.6
24:00 246.3 F 12.8
26:00 244.6 F 12.2
28:00 256.9 F 18.7
30:00 267.8 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 272.6 F 28.4
2:00 271.8 F 28.1
3:00 272.4 F 28.3
4:00 271.9 F 28.2

Min 271.5 F 27.7
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.0 F 0.0
5:00 194.3 F 0.0
10:00 196.4 F 0.0
15:00 197.0 F 0.0
20:00 200.6 F 0.0
25:00 190.9 F 0.0
30:00 209.9 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:21
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:18
CYCLE COMPLETE

09/30/2024 01:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____