

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2740

Name: RITTER 1  
Sterilizer ID: V2178198  
01/05/2023 10:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.5 F 0.0  
2:00 176.0 F 0.0  
4:00 185.6 F 0.5  
6:00 204.4 F 4.8  
8:00 217.1 F 9.5  
10:00 230.1 F 8.9  
12:00 234.9 F 9.9  
14:00 239.6 F 8.8  
16:00 245.2 F 13.1  
18:00 243.1 F 11.8  
20:00 255.8 F 18.4  
22:00 267.5 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.6 F 29.0  
2:00 271.5 F 28.2  
3:00 272.8 F 29.1  
4:00 271.8 F 28.3

Min 271.4 F 28.2  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.7 F 0.0  
5:00 217.8 F 0.0  
10:00 235.5 F 0.0  
15:00 230.6 F 0.0  
20:00 224.8 F 0.0  
25:00 225.6 F 0.0  
30:00 226.6 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 23:23  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:01:20  
CYCLE COMPLETE

01/05/2023 11:15 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_