

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3942

Name: RITTER 1  
Sterilizer ID: V2178198  
10/24/2023 11:06 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.0 F 0.0  
2:00 135.9 F 1.8  
4:00 150.8 F 4.6  
6:00 183.5 F 8.7  
8:00 216.7 F 9.6  
10:00 220.5 F 7.1  
12:00 233.5 F 11.8  
14:00 234.8 F 8.1  
16:00 243.1 F 12.2  
18:00 244.3 F 12.1  
20:00 256.6 F 18.7  
22:00 268.1 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 271.8 F 28.3  
2:00 272.4 F 28.5  
3:00 272.6 F 28.8  
4:00 272.4 F 29.0

Min 271.4 F 28.3  
Max 272.9 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.3 F 0.0  
5:00 231.5 F 0.0  
10:00 230.7 F 0.0  
15:00 229.9 F 0.0  
20:00 228.6 F 0.0  
25:00 219.2 F 0.0  
30:00 194.7 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 23:51  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:02:16  
CYCLE COMPLETE

10/24/2023 12:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_