

-----  
-----  
Midmark M11 - v1.0.5  
Total Cycles: 3995

Name: RITTER2  
Sterilizer ID: V2313228  
10/03/2024 01:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 138.6 F 0.0  
2:00 143.8 F 1.7  
4:00 158.8 F 4.8  
6:00 181.9 F 8.3  
8:00 195.6 F 12.2  
10:00 214.3 F 5.2  
12:00 223.9 F 9.1  
14:00 234.3 F 13.2  
16:00 235.0 F 9.8  
18:00 238.2 F 10.1  
20:00 244.9 F 13.3  
22:00 255.7 F 19.1  
24:00 265.6 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.7 F 29.9  
2:00 271.7 F 29.2  
3:00 271.7 F 29.7  
4:00 272.3 F 29.8

Min 271.2 F 29.1  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.4 F 0.0  
5:00 199.0 F 0.0  
10:00 197.4 F 0.0  
15:00 193.0 F 0.0  
20:00 190.9 F 0.0  
25:00 190.2 F 0.0  
30:00 191.9 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 25:54  
STERILIZING: 4:00  
VENTING: 3:51  
DRYING: 30:00  
TOTAL CYCLE: 01:04:50  
CYCLE COMPLETE

10/03/2024 02:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_