
Midmark M11 - v1.0.0
Total Cycles: 1414

Name: RITTER 1
Sterilizer ID: V2178198
12/01/2021 10:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.0 F 0.0
2:00 159.4 F 1.9
4:00 171.3 F 5.4
6:00 200.5 F 10.2
8:00 219.8 F 5.0
10:00 229.7 F 9.2
12:00 238.1 F 10.1
14:00 241.0 F 11.4
16:00 237.9 F 9.4
18:00 251.7 F 16.1
20:00 263.3 F 22.8
22:00 272.0 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.7 F 29.1
2:00 271.7 F 28.4
3:00 272.6 F 29.0
4:00 271.8 F 28.8

Min 270.9 F 28.1
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.5 F 0.0
5:00 226.9 F 0.0
10:00 218.3 F 0.0
15:00 227.2 F 0.0
20:00 202.8 F 0.0
25:00 221.3 F 0.0
30:00 213.4 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 22:13
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:44
CYCLE COMPLETE

12/01/2021 11:42 AM
ACCEPTED: X
REJECTED:
Operator: U02