
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001084
9:17 12/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:52
PRESSURIZING 2:58
STERILIZING 4:20
136.4 C 220kPa 4:20
136.3 C 224kPa 4:50
136.4 C 223kPa 5:20
136.3 C 221kPa 5:50
136.4 C 224kPa 6:20
136.3 C 221kPa 6:50
136.4 C 224kPa 7:20
136.3 C 223kPa 7:50
136.3 C 223kPa 7:50
VENTING 7:50
AIR DRYING 8:44
CYCLE COMPLETE 8:46

Digital Signature #
E9C749E0FDE7507A
