
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006618
9:34 15/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:37
STERILIZING 3:39
136.5 C 223kPa 3:39
135.8 C 220kPa 4:09
135.9 C 220kPa 4:39
136.4 C 223kPa 5:09
136.3 C 224kPa 5:39
136.3 C 224kPa 6:09
136.4 C 223kPa 6:39
136.4 C 224kPa 7:09
136.4 C 224kPa 7:09
VENTING 7:09
AIR DRYING 8:00
CYCLE COMPLETE 8:05

Digital Signature #
6B5EEC6CD4380946
