
Midmark M11 - v1.0.5
Total Cycles: 633

Name: RITTER2
Sterilizer ID: V2313228
12/15/2021 07:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.4 F 0.0
2:00 92.7 F 0.0
4:00 77.5 F 0.1
6:00 108.7 F 1.2
8:00 144.3 F 3.8
10:00 167.5 F 7.3
12:00 188.9 F 11.3
14:00 211.8 F 3.5
16:00 221.7 F 7.9
18:00 233.3 F 12.2
20:00 233.6 F 8.4
22:00 244.0 F 12.5
24:00 243.5 F 12.2
26:00 255.4 F 18.4
28:00 266.1 F 24.9
30:00 273.1 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.7 F 29.0
2:00 272.5 F 29.4
3:00 272.0 F 29.4
4:00 272.2 F 28.9

Min 271.6 F 28.8
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.0 F 0.0
5:00 218.7 F 0.0
10:00 217.5 F 0.0
15:00 196.2 F 0.0
20:00 188.1 F 0.0
25:00 221.0 F 0.0
30:00 238.4 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 30:20
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 01:08:24
CYCLE COMPLETE

12/15/2021 09:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____