
Midmark M11 - v1.0.5
Total Cycles: 4404

Name: RITTER2
Sterilizer ID: V2313228
03/05/2025 05:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.4 F 0.0
2:00 128.7 F 1.4
4:00 135.9 F 4.2
6:00 168.8 F 7.8
8:00 185.3 F 11.7
10:00 214.1 F 5.4
12:00 225.5 F 9.5
14:00 232.7 F 11.1
16:00 236.2 F 11.0
18:00 234.7 F 8.9
20:00 247.1 F 14.9
22:00 257.7 F 20.8
24:00 267.8 F 27.3
26:00 272.4 F 31.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 30.6
1:00 272.4 F 30.7
2:00 272.2 F 31.1
3:00 271.7 F 30.6
4:00 272.4 F 30.9

Min 271.7 F 30.5
Max 272.5 F 31.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.1 F 0.0
5:00 226.0 F 0.0
10:00 213.4 F 0.0
15:00 205.2 F 0.0
20:00 199.8 F 0.0
25:00 201.6 F 0.0
30:00 200.9 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 26:27
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:05:30
CYCLE COMPLETE

03/05/2025 06:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____