

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4348

Name: RITTER 1  
Sterilizer ID: V2178198  
02/29/2024 03:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.8 F 0.0  
2:00 172.0 F 3.5  
4:00 210.8 F 7.8  
6:00 208.2 F 13.1  
8:00 224.7 F 6.5  
10:00 234.5 F 11.5  
12:00 234.9 F 8.3  
14:00 241.8 F 11.1  
16:00 245.7 F 12.9  
18:00 258.3 F 19.6  
20:00 269.5 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 28.3  
1:00 272.6 F 28.6  
2:00 272.7 F 28.9  
3:00 271.6 F 28.3  
4:00 272.6 F 28.6  
  
Min 271.6 F 28.2  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.7 F 0.0  
5:00 230.9 F 0.0  
10:00 225.0 F 0.0  
15:00 233.1 F 0.0  
20:00 226.3 F 0.0  
25:00 229.1 F 0.0  
30:00 224.4 F 0.0  
Drying Complete

FILLING: 1:20  
HEATING: 21:33  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:00:17  
CYCLE COMPLETE

02/29/2024 04:19 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_