
Midmark M11 - v1.0.5
Total Cycles: 3016

Name: RITTER2
Sterilizer ID: V2313228
10/17/2023 08:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 81.7 F 0.0
4:00 77.3 F 0.4
6:00 112.4 F 2.5
8:00 142.9 F 5.3
10:00 171.0 F 8.6
12:00 191.7 F 12.9
14:00 205.2 F 5.4
16:00 219.3 F 9.3
18:00 231.0 F 11.8
20:00 234.9 F 10.1
22:00 235.3 F 8.7
24:00 246.6 F 14.0
26:00 258.5 F 20.4
28:00 269.0 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.6
1:00 272.0 F 29.5
2:00 272.0 F 29.5
3:00 272.0 F 29.4
4:00 272.2 F 29.5

Min 271.7 F 29.4
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.1 F 0.0
5:00 210.3 F 0.0
10:00 190.8 F 0.0
15:00 191.1 F 0.0
20:00 188.8 F 0.0
25:00 187.4 F 0.0
30:00 185.6 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 29:46
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:07:45
CYCLE COMPLETE

10/17/2023 09:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____