

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4708

Name: RITTER 1  
Sterilizer ID: V2178198  
07/10/2024 09:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.1 F 0.0  
2:00 74.4 F 0.1  
4:00 78.1 F 1.0  
6:00 97.3 F 2.7  
8:00 126.5 F 5.1  
10:00 153.4 F 7.9  
12:00 174.2 F 11.0  
14:00 207.8 F 5.6  
16:00 210.7 F 4.7  
18:00 219.2 F 7.8  
20:00 228.5 F 11.3  
22:00 231.9 F 6.6  
24:00 239.4 F 11.3  
26:00 236.3 F 8.6  
28:00 248.4 F 14.5  
30:00 258.6 F 20.1  
32:00 268.3 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 271.7 F 29.0  
2:00 272.7 F 29.3  
3:00 271.6 F 29.0  
4:00 272.7 F 29.4  
  
Min 270.9 F 28.3  
Max 273.1 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.2 F 0.0  
5:00 208.2 F 0.0  
10:00 206.1 F 0.0  
15:00 230.4 F 0.0  
20:00 231.8 F 0.0  
25:00 231.5 F 0.0  
30:00 229.6 F 0.0  
  
Drying Complete

FILLING: 0:48  
HEATING: 33:17  
STERILIZING: 4:00  
VENTING: 3:51  
DRYING: 30:00  
TOTAL CYCLE: 01:11:57  
CYCLE COMPLETE

07/10/2024 10:32 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_