
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002554
15:07 28/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:57
STERILIZING 2:57
136.4 C 223kPa 2:57
136.1 C 223kPa 3:27
136.3 C 225kPa 3:57
136.4 C 224kPa 4:27
136.4 C 227kPa 4:57
136.4 C 227kPa 5:27
136.4 C 226kPa 5:57
136.4 C 227kPa 6:27
136.4 C 227kPa 6:27
VENTING 6:27
AIR DRYING 7:21
CYCLE COMPLETE 18:54

Digital Signature #
A69400176FCBA23A
