
Midmark M11 - v1.0.0
Total Cycles: 4722

Name: RITTER 1
Sterilizer ID: V2178198
07/12/2024 09:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.8 F 0.0
2:00 139.1 F 1.8
4:00 144.4 F 4.5
6:00 167.7 F 7.9
8:00 188.8 F 11.8
10:00 209.7 F 4.1
12:00 219.0 F 8.5
14:00 230.7 F 12.9
16:00 233.2 F 8.8
18:00 241.4 F 11.6
20:00 242.8 F 11.9
22:00 254.2 F 17.6
24:00 264.3 F 23.6
26:00 272.0 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.3 F 29.3
2:00 272.3 F 29.0
3:00 271.5 F 28.5
4:00 271.2 F 28.7

Min 271.1 F 28.3
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.9 F 0.0
5:00 203.0 F 0.0
10:00 208.5 F 0.0
15:00 219.6 F 0.0
20:00 226.3 F 0.0
25:00 232.3 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 26:14
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:04:49
CYCLE COMPLETE

07/12/2024 10:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____