

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001246  
17:14 02/09/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:30  
PRESSURIZING 4:31  
STERILIZING 5:52  
136.4 C 224kPa 5:52  
136.3 C 223kPa 6:22  
136.4 C 223kPa 6:52  
136.1 C 221kPa 7:22  
136.2 C 222kPa 7:52  
136.4 C 225kPa 8:22  
136.2 C 223kPa 8:52  
136.2 C 223kPa 9:22  
136.2 C 223kPa 9:22  
VENTING 9:22  
AIR DRYING 10:20  
CYCLE COMPLETE 19:40  
-----

Digital Signature #  
6517A1F0BA9A93C4  
-----