
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001757
10:23 01/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:28
STERILIZING 3:17
136.0 C 220kPa 3:17
136.4 C 224kPa 3:47
136.4 C 225kPa 4:17
136.4 C 226kPa 4:47
136.2 C 223kPa 5:17
136.4 C 223kPa 5:47
136.4 C 221kPa 6:17
136.0 C 222kPa 6:47
136.0 C 222kPa 6:47
VENTING 6:47
AIR DRYING 7:35
CYCLE COMPLETE 8:03

Digital Signature #
28EB82C9CE27D711
