

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2215

Name: RITTER2  
Sterilizer ID: V2313228  
04/04/2023 12:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.8 F 0.0  
2:00 184.1 F 3.8  
4:00 194.1 F 8.7  
6:00 217.2 F 8.1  
8:00 229.9 F 9.6  
10:00 234.2 F 8.9  
12:00 243.4 F 12.4  
14:00 245.4 F 13.4  
16:00 258.8 F 20.6  
18:00 270.5 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.6 F 29.6  
2:00 271.5 F 29.0  
3:00 272.6 F 29.7  
4:00 272.1 F 29.1  
  
Min 271.5 F 28.8  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.2 F 0.0  
5:00 237.9 F 0.0  
10:00 229.8 F 0.0  
15:00 228.7 F 0.0  
20:00 226.9 F 0.0  
25:00 224.8 F 0.0  
30:00 222.2 F 0.0  
Drying Complete

FILLING: 1:57  
HEATING: 18:47  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:57:56  
CYCLE COMPLETE

04/04/2023 01:11 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_