
Midmark M11 - v1.0.0
Total Cycles: 1828

Name: RITTER 1
Sterilizer ID: V2178198
04/14/2022 03:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.0 F 0.0
2:00 194.3 F 4.0
4:00 206.3 F 8.8
6:00 222.5 F 7.2
8:00 230.5 F 9.1
10:00 236.4 F 9.7
12:00 242.4 F 12.1
14:00 241.0 F 10.8
16:00 254.8 F 17.8
18:00 266.7 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.4 F 28.8
2:00 271.3 F 28.3
3:00 272.4 F 28.9
4:00 271.6 F 28.3

Min 271.2 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 208.9 F 0.0
10:00 207.5 F 0.0
15:00 213.3 F 0.0
20:00 218.6 F 0.0
25:00 222.2 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 19:29
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:48
CYCLE COMPLETE

04/14/2022 04:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____