
Midmark M11 - v1.0.5
Total Cycles: 1417

Name: RITTER2
Sterilizer ID: V2313228
08/25/2022 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.7 F 0.0
2:00 184.5 F 3.1
4:00 188.1 F 7.4
6:00 208.0 F 12.7
8:00 227.3 F 8.5
10:00 240.7 F 11.2
12:00 241.9 F 12.3
14:00 241.9 F 11.3
16:00 254.7 F 18.0
18:00 266.1 F 24.8
20:00 272.4 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.3
1:00 272.8 F 29.4
2:00 272.8 F 29.9
3:00 271.9 F 29.2
4:00 272.6 F 29.7

Min 271.4 F 29.1
Max 273.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.8 F 0.0
5:00 230.1 F 0.0
10:00 242.6 F 0.0
15:00 233.6 F 0.0
20:00 224.0 F 0.0
25:00 222.8 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 20:27
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:18
CYCLE COMPLETE

08/25/2022 02:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____