

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5056

Name: RITTER 1  
Sterilizer ID: V2178198  
11/05/2024 10:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.9 F 0.0  
2:00 168.7 F 0.4  
4:00 173.7 F 2.3  
6:00 212.6 F 6.3  
8:00 212.5 F 10.7  
10:00 226.3 F 5.8  
12:00 228.8 F 9.0  
14:00 240.2 F 12.5  
16:00 240.7 F 10.9  
18:00 236.2 F 8.7  
20:00 249.6 F 14.9  
22:00 259.9 F 21.0  
24:00 270.1 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.7 F 29.2  
2:00 272.5 F 28.7  
3:00 271.4 F 28.4  
4:00 272.4 F 29.2

Min 271.3 F 28.3  
Max 273.0 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.5 F 0.0  
5:00 224.2 F 0.0  
10:00 194.1 F 0.0  
15:00 192.7 F 0.0  
20:00 189.9 F 0.0  
25:00 187.6 F 0.0  
30:00 188.2 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 24:60  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:03:39  
CYCLE COMPLETE

11/05/2024 12:01 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_