

-----  
Midmark M11 - v1.0.5  
Total Cycles: 729

Name: RITTER2  
Sterilizer ID: V2313228  
01/26/2022 09:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 102.8 F 0.0  
2:00 105.5 F 0.7  
4:00 118.1 F 2.7  
6:00 151.0 F 5.7  
8:00 175.9 F 9.3  
10:00 196.3 F 13.4  
12:00 208.3 F 6.4  
14:00 222.6 F 11.0  
16:00 227.1 F 6.8  
18:00 239.7 F 12.4  
20:00 238.8 F 10.3  
22:00 251.0 F 16.3  
24:00 261.6 F 22.4  
26:00 271.4 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.2 F 29.9  
2:00 272.6 F 29.6  
3:00 271.7 F 29.1  
4:00 271.3 F 29.1

Min 271.2 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.7 F 0.0  
5:00 215.8 F 0.0  
10:00 188.1 F 0.0  
15:00 184.3 F 0.0  
20:00 179.9 F 0.0  
25:00 179.0 F 0.0  
30:00 177.8 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 26:45  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:04:50  
CYCLE COMPLETE

01/26/2022 10:45 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_