
Midmark M11 - v1.0.5
Total Cycles: 1700

Name: RITTER2
Sterilizer ID: V2313228
11/08/2022 03:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.2 F 0.0
2:00 173.7 F 0.5
4:00 196.4 F 4.2
6:00 218.4 F 9.4
8:00 230.6 F 8.4
10:00 237.5 F 11.6
12:00 237.6 F 9.9
14:00 237.7 F 8.7
16:00 251.5 F 16.2
18:00 264.0 F 23.5
20:00 271.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.5 F 29.1
2:00 271.6 F 29.0
3:00 272.4 F 29.1
4:00 271.4 F 28.8

Min 271.4 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 227.9 F 0.0
10:00 231.0 F 0.0
15:00 226.7 F 0.0
20:00 223.2 F 0.0
25:00 221.2 F 0.0
30:00 221.2 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 20:01
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

11/08/2022 04:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____