

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3985

Name: RITTER2  
Sterilizer ID: V2313228  
10/01/2024 03:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.1 F 0.0  
2:00 159.8 F 2.8  
4:00 170.8 F 6.2  
6:00 192.9 F 10.4  
8:00 210.4 F 6.0  
10:00 220.3 F 8.1  
12:00 230.8 F 12.7  
14:00 234.8 F 9.8  
16:00 237.9 F 9.6  
18:00 247.3 F 14.5  
20:00 258.7 F 20.7  
22:00 268.9 F 27.3  
24:00 271.9 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.6  
1:00 272.6 F 30.0  
2:00 272.0 F 30.0  
3:00 272.4 F 29.8  
4:00 272.1 F 30.1

Min 271.6 F 29.5  
Max 272.7 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.9 F 0.0  
5:00 215.4 F 0.0  
10:00 213.2 F 0.0  
15:00 210.5 F 0.0  
20:00 214.0 F 0.0  
25:00 213.0 F 0.0  
30:00 210.8 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 24:03  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:03:07  
CYCLE COMPLETE

10/01/2024 04:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_