
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 004752
11:55 22/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:29
STERILIZING 3:26
136.4 C 223kPa 3:26
136.1 C 219kPa 3:56
136.2 C 221kPa 4:26
136.2 C 223kPa 4:56
136.4 C 223kPa 5:26
136.4 C 222kPa 5:56
136.3 C 223kPa 6:26
136.4 C 223kPa 6:56
136.4 C 223kPa 6:56
VENTING 6:56
AIR DRYING 7:47
CYCLE COMPLETE 8:00

Digital Signature #
E4CAD0E247A25534
