
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002373
10:20 13/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:17
PRESSURIZING 2:26
STERILIZING 3:41
136.4 C 225kPa 3:41
136.3 C 224kPa 4:11
136.0 C 222kPa 4:41
136.4 C 225kPa 5:11
136.4 C 225kPa 5:41
136.4 C 224kPa 6:11
136.3 C 224kPa 6:41
136.3 C 223kPa 7:11
136.3 C 223kPa 7:11
VENTING 7:11
AIR DRYING 8:02
CYCLE COMPLETE 11:03

Digital Signature #
90171A5C63B8DAB9
