

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001634  
14:32 29/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:15  
PRESSURIZING 2:22  
STERILIZING 3:25  
136.0 C 222kPa 3:25  
136.3 C 225kPa 3:55  
136.3 C 225kPa 4:25  
136.3 C 225kPa 4:55  
136.0 C 222kPa 5:25  
136.4 C 225kPa 5:55  
136.4 C 226kPa 6:25  
136.4 C 226kPa 6:55  
136.4 C 226kPa 6:55  
VENTING 6:55  
AIR DRYING 7:45  
CYCLE COMPLETE 7:50  
-----

Digital Signature #  
BD6F10AAF0CAE251  
-----