
Midmark M11 - v1.0.0
Total Cycles: 1079

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2021 11:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.5 F 0.0
2:00 141.4 F 1.0
4:00 164.9 F 3.2
6:00 187.6 F 6.4
8:00 204.6 F 11.6
10:00 215.4 F 4.4
12:00 227.3 F 8.8
14:00 238.3 F 13.4
16:00 238.9 F 10.5
18:00 236.0 F 8.4
20:00 249.7 F 15.1
22:00 261.0 F 21.4
24:00 271.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.5 F 28.9
2:00 271.5 F 28.2
3:00 272.6 F 29.1
4:00 272.0 F 28.5

Min 271.5 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 213.0 F 0.0
10:00 219.7 F 0.0
15:00 218.9 F 0.0
20:00 222.1 F 0.0
25:00 222.3 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 24:40
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:03:05
CYCLE COMPLETE

08/09/2021 12:35 PM
ACCEPTED: X
REJECTED:
Operator: U01