

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2857

Name: RITTER2  
Sterilizer ID: V2313228  
09/11/2023 08:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.1 F 0.0  
2:00 71.8 F 0.0  
4:00 105.8 F 0.0  
6:00 95.7 F 0.0  
8:00 147.6 F 0.0  
10:00 172.7 F 2.3  
12:00 193.5 F 5.5  
14:00 210.5 F 8.9  
16:00 219.1 F 12.5  
18:00 224.9 F 6.6  
20:00 234.4 F 10.7  
22:00 236.5 F 8.2  
24:00 241.5 F 12.2  
26:00 239.5 F 10.4  
28:00 251.7 F 16.6  
30:00 262.5 F 22.8  
32:00 272.7 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.4 F 29.2  
2:00 271.9 F 29.4  
3:00 271.9 F 28.9  
4:00 272.5 F 29.6  
  
Min 271.3 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.4 F 0.0  
5:00 198.6 F 0.0  
10:00 196.7 F 0.0  
15:00 190.8 F 0.0  
20:00 196.1 F 0.0  
25:00 198.3 F 0.0  
30:00 198.4 F 0.0  
  
Drying Complete

FILLING: 0:48  
HEATING: 32:21  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:10:25  
CYCLE COMPLETE

09/11/2023 09:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_