
Midmark M11 - v1.0.0
Total Cycles: 3061

Name: RITTER 1
Sterilizer ID: V2178198
03/30/2023 02:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.0 F 0.0
2:00 131.7 F 0.2
4:00 157.9 F 3.3
6:00 183.6 F 7.4
8:00 209.3 F 12.5
10:00 219.6 F 6.4
12:00 232.3 F 11.4
14:00 232.7 F 8.0
16:00 242.0 F 11.5
18:00 245.3 F 12.7
20:00 257.9 F 19.5
22:00 269.2 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.2 F 28.8
2:00 271.3 F 28.4
3:00 272.3 F 28.9
4:00 271.2 F 28.2

Min 271.2 F 28.2
Max 272.4 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.2 F 0.0
5:00 210.2 F 0.0
10:00 223.9 F 0.0
15:00 242.4 F 0.0
20:00 232.7 F 0.0
25:00 228.8 F 0.0
30:00 224.9 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 23:01
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:01:12
CYCLE COMPLETE

03/30/2023 03:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____