
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001439
9:09 07/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:49
STERILIZING 3:50
136.5 C 224kPa 3:50
136.4 C 225kPa 4:20
136.2 C 223kPa 4:50
136.4 C 223kPa 5:20
136.4 C 223kPa 5:50
136.2 C 223kPa 6:20
136.1 C 222kPa 6:50
136.1 C 222kPa 7:20
136.1 C 222kPa 7:20
VENTING 7:20
AIR DRYING 8:09
CYCLE COMPLETE 8:16

Digital Signature #
5E11172E8FF82DD5
