
Midmark M11 - v1.0.0
Total Cycles: 2797

Name: RITTER 1
Sterilizer ID: V2178198
01/19/2023 11:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.7 F 0.0
2:00 196.8 F 3.9
4:00 197.3 F 8.6
6:00 214.8 F 9.9
8:00 225.3 F 8.1
10:00 238.3 F 13.5
12:00 239.8 F 11.0
14:00 238.1 F 9.5
16:00 252.0 F 16.4
18:00 264.0 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 29.1
2:00 271.6 F 28.3
3:00 272.4 F 29.1
4:00 272.3 F 28.8

Min 271.4 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 237.8 F 0.0
10:00 229.9 F 0.0
15:00 229.2 F 0.0
20:00 227.9 F 0.0
25:00 228.8 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 19:59
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:23
CYCLE COMPLETE

01/19/2023 12:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____