

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3832

Name: RITTER2  
Sterilizer ID: V2313228  
08/07/2024 09:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.5 F 0.0  
2:00 156.9 F 1.8  
4:00 165.7 F 4.7  
6:00 189.0 F 8.8  
8:00 206.5 F 13.3  
10:00 223.6 F 6.9  
12:00 230.4 F 11.8  
14:00 233.8 F 8.8  
16:00 240.3 F 11.0  
18:00 245.8 F 13.5  
20:00 257.3 F 19.7  
22:00 268.0 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.9  
1:00 272.1 F 29.4  
2:00 271.3 F 29.2  
3:00 272.0 F 29.3  
4:00 271.3 F 29.3

Min 270.9 F 28.8  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.9 F 0.0  
5:00 231.6 F 0.0  
10:00 219.1 F 0.0  
15:00 184.8 F 0.0  
20:00 180.5 F 0.0  
25:00 179.9 F 0.0  
30:00 179.7 F 0.0  
Drying Complete

FILLING: 1:16  
HEATING: 23:21  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:02:12  
CYCLE COMPLETE

08/07/2024 10:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_