
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001984
9:04 03/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:32
PRESSURIZING 3:35
STERILIZING 4:41
136.0 C 220kPa 4:41
136.0 C 221kPa 5:11
CYCLE ABORTED 5:24
STOP BUTTON PRESSED
NOT STERILE
END TIME 5:24

Digital Signature #
B67534288F233B11
