
Midmark M11 - v1.0.5
Total Cycles: 4217

Name: RITTER2
Sterilizer ID: V2313228
12/16/2024 11:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	167.0 F	0.0
2:00	161.1 F	1.6
4:00	159.4 F	4.7
6:00	175.8 F	8.5
8:00	196.4 F	13.2
10:00	216.0 F	7.0
12:00	227.8 F	11.8
14:00	231.0 F	8.5
16:00	241.0 F	11.6
18:00	243.7 F	12.8
20:00	255.4 F	19.0
22:00	266.1 F	25.5
24:00	272.5 F	30.1

STERILIZING

mm:ss	Degrees	PSI
0:00	271.9 F	29.7
1:00	272.4 F	29.7
2:00	272.7 F	30.0
3:00	272.6 F	30.2
4:00	272.0 F	29.7

Min 271.9 F 29.6
Max 272.9 F 30.2

VENTING CHAMBER

DRYING

Drying Start

mm:ss	Degrees	PSI
0:00	209.6 F	0.0
5:00	209.0 F	0.0
10:00	207.8 F	0.0
15:00	194.9 F	0.0
20:00	199.1 F	0.0
25:00	203.9 F	0.0
30:00	205.9 F	0.0

Drying Complete

FILLING: 1:03
HEATING: 24:26
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:03:14
CYCLE COMPLETE

12/16/2024 12:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____