
Midmark M11 - v1.0.5
Total Cycles: 469

Name: RITTER2
Sterilizer ID: V2313228
10/21/2021 11:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.6 F 0.0
2:00 184.3 F 3.8
4:00 200.5 F 9.3
6:00 220.9 F 7.3
8:00 235.5 F 11.4
10:00 235.0 F 9.5
12:00 237.2 F 9.3
14:00 249.8 F 15.7
16:00 262.1 F 22.9
18:00 272.5 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.2 F 29.1
2:00 272.4 F 29.5
3:00 272.1 F 29.6
4:00 272.0 F 29.1

Min 271.5 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.8 F 0.0
5:00 223.4 F 0.0
10:00 226.7 F 0.0
15:00 218.8 F 0.0
20:00 212.0 F 0.0
25:00 218.8 F 0.0
30:00 224.1 F 0.0

Drying Complete

FILLING: 2:24
HEATING: 18:09
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:57:37
CYCLE COMPLETE

10/21/2021 12:30 PM
ACCEPTED: X
REJECTED:
Operator: U02