
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001911
8:14 31/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:55
STERILIZING 3:48
136.0 C 220kPa 3:48
136.4 C 224kPa 4:18
136.4 C 224kPa 4:48
136.4 C 224kPa 5:18
136.4 C 224kPa 5:48
135.9 C 220kPa 6:18
136.3 C 222kPa 6:48
136.4 C 224kPa 7:18
136.4 C 224kPa 7:18
VENTING 7:18
AIR DRYING 8:05
CYCLE COMPLETE 38:05

Digital Signature #
5E0A9E5E0759D707
