

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005868  
11:04 29/06/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:00  
PRESSURIZING 2:01  
STERILIZING 3:04  
136.4 C 224kPa 3:04  
136.4 C 222kPa 3:34  
136.4 C 225kPa 4:04  
136.4 C 222kPa 4:34  
136.4 C 222kPa 5:04  
136.4 C 219kPa 5:34  
136.4 C 221kPa 6:04  
136.4 C 225kPa 6:34  
136.4 C 225kPa 6:34  
VENTING 6:35  
AIR DRYING 7:25  
CYCLE COMPLETE 7:30

-----  
Digital Signature #  
A3F6BAEC3EC16662  
-----