
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006366
8:22 27/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 2:59
STERILIZING 3:54
136.0 C 220kPa 3:54
136.1 C 223kPa 4:24
136.2 C 222kPa 4:54
136.2 C 220kPa 5:24
136.0 C 220kPa 5:54
136.1 C 222kPa 6:24
136.4 C 223kPa 6:54
136.2 C 219kPa 7:24
136.2 C 219kPa 7:24
VENTING 7:24
AIR DRYING 8:15
CYCLE COMPLETE 27:06

Digital Signature #
BC5166D19E2BF94D
