

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3229

Name: RITTER2  
Sterilizer ID: V2313228  
12/18/2023 02:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 125.3 F 0.0  
2:00 124.6 F 0.8  
4:00 151.8 F 3.3  
6:00 170.2 F 7.0  
8:00 193.3 F 11.6  
10:00 216.7 F 5.5  
12:00 222.5 F 10.2  
14:00 231.6 F 8.3  
16:00 240.2 F 12.0  
18:00 238.3 F 10.3  
20:00 251.3 F 16.6  
22:00 262.6 F 23.2  
24:00 272.9 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.2  
1:00 272.8 F 30.3  
2:00 271.9 F 29.9  
3:00 272.0 F 29.5  
4:00 272.7 F 29.9

Min 271.0 F 29.0  
Max 273.0 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.3 F 0.0  
5:00 241.3 F 0.0  
10:00 210.9 F 0.0  
15:00 216.8 F 0.0  
20:00 213.4 F 0.0  
25:00 218.2 F 0.0  
30:00 206.8 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 24:22  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:02:36  
CYCLE COMPLETE

12/18/2023 03:42 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_