

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000632  
13:45 03/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:40  
STERILIZING 3:47  
136.4 C 220kPa 3:47  
136.1 C 231kPa 4:17  
136.2 C 230kPa 4:47  
136.3 C 231kPa 5:17  
136.4 C 232kPa 5:47  
136.3 C 232kPa 6:17  
136.5 C 233kPa 6:47  
136.3 C 232kPa 7:17  
136.3 C 232kPa 7:17  
VENTING 7:17  
AIR DRYING 8:10  
CYCLE COMPLETE 23:10  
-----

Digital Signature #  
D6363B4E3CB8A7D0  
-----