
Midmark M11 - v1.0.5
Total Cycles: 2805

Name: RITTER2
Sterilizer ID: V2313228
08/28/2023 02:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.7 F 0.0
2:00 127.0 F 0.0
4:00 155.0 F 0.0
6:00 185.9 F 0.8
8:00 205.1 F 4.3
10:00 216.6 F 8.0
12:00 228.9 F 12.3
14:00 230.7 F 8.3
16:00 241.6 F 13.4
18:00 241.1 F 11.5
20:00 239.7 F 10.4
22:00 252.6 F 16.9
24:00 263.7 F 23.4
26:00 272.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.6 F 29.1
2:00 272.1 F 29.3
3:00 272.1 F 28.8
4:00 272.5 F 29.5

Min 271.5 F 28.6
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 206.0 F 0.0
10:00 207.9 F 0.0
15:00 202.4 F 0.0
20:00 202.1 F 0.0
25:00 202.7 F 0.0
30:00 201.4 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 26:08
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:04:27
CYCLE COMPLETE

08/28/2023 03:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____